



Fabrication and application of 3D force sensors based on piezoelectric and piezoresistive effects

János Radó

MTA EK MFA MEMS laboratory

Fourth semester

Supervisor: Csikósné Papp Andrea Edit

Introduction

3 parts of my work

- Piezoresistive 3D force sensors***
- Piezoelectric ZnO nano-rods***
- Piezoelectric thin films (for 3D force sensor)***

Piezoresistive 3D force sensors

Applications

- Integration in a surgery robot's laparoscope***
- Integration in a vehicle tyre***

Piezoresistive 3D force sensors - Integration in a laparoscopic tool

3D force sensor chip



Specific requirements:

- Reduced size (down to $1 \times 1 \text{ mm}^2$) to enable integration
- Sensitivity: 1-20N for gripping force, 10-1000mN for tactile sensing
- Robustness (vs. sensitivity)
- Biocompatible coating can withstand sterilization

Operation:

- Deforming c-Si membrane
- 4 embedded piezoresistors
- 4 Voltage dividers or Wheatstone bridges
- Calculation of vectorial components:



$$F_x = \frac{1}{V_0 \alpha_{ls} \pi_{44}} (\Delta V_{right} - \Delta V_{left}),$$
$$F_y = \frac{1}{V_0 \alpha_{ls} \pi_{44}} (\Delta V_{top} - \Delta V_{bottom}),$$
$$F_z = \frac{1}{V_0 \alpha_{ls} \pi_{44}} \frac{(\Delta V_{left} + \Delta V_{right} + \Delta V_{top} + \Delta V_{bottom})}{2}$$





Piezoresistive 3D force sensors - Integration in a laparoscopic tool

Previous work

- Fabrication



- Testing



Piezoresistive 3D force sensors - Integration in a laparoscopic tool










Previous work

- Measurements

Effect of elastic coating



Bare sensor

Covered sensor




Piezoresistive 3D force sensors - Integration in a laparoscopic tool

Previous work – metal gripper and biocompatible coating (Nusil MED-6215)




Piezoresistive 3D force sensors - Integration in a laparoscopic tool

Previous work – temperature tests



Temperature profile




Offset signals




Climate chamber

Piezoresistive 3D force sensors - Integration in a laparoscopic tool

Previous work – response time



Bare sensor




Covered sensor






Piezoresistive 3D force sensors - Integration in a laparoscopic tool

Previous work – completed test device




Piezoresistive 3D force sensors - Integration in a laparoscopic tool

Current work – biomechanical tests



80 – 120Hgmm
106 – 160mBar





Piezoresistive 3D force sensors - Integration in a laparoscopic tool

Current work – biomechanical tests




Piezoresistive 3D force sensors - Integration in a laparoscopic tool

Current work – biomechanical tests



Piezoresistive 3D force sensors - Integration in a laparoscopic tool

Current work – biomechanical tests





Piezoresistive 3D force sensors – Reducing chip size

(Supported by the ÚNKP-17-3-I-OE-779/47 New National Excellence Program of the Ministry Of Human Capacities)

Means:







- Replacement of glass substrate
- Application of new contacts
- Design new read-out electronics



Piezoresistive 3D force sensors - Reducing chip size

(Supported by the ÚNKP-17-3-I-OE-779/47 New National Excellence Program of the Ministry Of Human Capacities)

UBM





Piezoresistive 3D force sensors - Reducing chip size

(Supported by the ÚNKP-17-3-I-OE-779/47 New National Excellence Program of the Ministry Of Human Capacities)

Wafer bonding

Aspects:

- Importance of cleanness of the surface
- Strong bond between SOI wafer and oxidized surface
- Hard structured geometry



Piezoresistive 3D force sensors - Integration in a laparoscopic tool

Publications in this topic

Papers:

- János Radó, Csaba Dücső, Gábor Battistig, Gábor Szebényi, Zbigniew Nawrat, Kamil Rohr, Péter Fürjes: **3D force sensors for laparoscopic surgery tool**, Microsystem Technologies, Volume 24., pp. 519-525, 2018
- K.Rohr, P.Fürjes, L.Mucha, J.Radó, K.Lis, Cs.Dücső, W.Sadowski, P.Földesy, D.Krawczyk, P.KroczeK, Z.Malota, G.Szebényi, Z.Nawrat, **Robin Heart Force Feedback/Control System**, Medical Robotics Reports, Volume 4, 2015, ISSN 2299-7407

Oral presentation (in English):

- J. Radó, G. Battistig, A.E. Pap, P. Fürjes, P. Földesy , **Thermal noise limited, scalable multi-piezoresistor readout architecture**, EUROSENSORS 2017, Paris, France
- J.Radó, Cs.Dücső, P.Fürjes, P.Földesy, G.Szebényi, Z.Nawrat, **3D force sensors for laparoscopic surgery tool/for surgery robotics**, ROBOTY MEDYCZNE 2016, Zabrze, Poland
- János Radó, Csaba Dücső, Gábor Battistig, Gábor Szebényi, Zbigniew Nawrat, Kamil Rohr, Péter Fürjes: **3D force sensors for laparoscopic surgery tool**, DTIP 2016, Budapest,
- Kamil Rohr, Lukasz Mucha, Péter Fürjes, János Radó, Csaba Dücső, Péter Földesy, Gábor Szebényi, Zbigniew Nawrat: **Robin Heart Force Feedback/Control**, MEDICAL ROBOTS 2015, Zabrze, Poland




Poster:

- Radó János, Dücső Csaba, Battistig Gábor, Fürjes Péter: **3D mikro-erőmérő sebészrobot alkalmazáshoz**, ORSZÁGOS ANYAGTUDOMÁNYI KONFERENCIA 2015, poszter szekció, magyar nyelvű

Piezoresistive 3D force sensors - Integration in a vehicle tyre

Previous work

- Implantation of 3D force sensor in a special rubber




- Integration of test tool in a vehicle tyre




Piezoresistive 3D force sensors - Integration in a vehicle tyre

Previous work

- Measurements – normal road conditions




- Measurements – the wheel is blocked



Piezoresistive 3D force sensors - Integration in a vehicle tyre

Previous work


- Functional test: static loading



Piezoresistive 3D force sensors - Integration in a vehicle tyre

Current work

- Results of static measurement




F_x —
 F_y —
 F_z —

Piezoresistive 3D force sensors - Integration in a vehicle tyre

Current work

- **Multiple regression:** demonstration of proportional relationship between tyre deformation and acting forces on the wheel



Piezoresistive 3D force sensors - Integration in a vehicle tyre

Current work

- Design new wireless readout circuit


Future work

- Building a special test station
- More tests in real circumstances

Publications in this topic

Oral presentation (in English):




- J. Radó, G. Battistig, S. Kuliniy, R. Végvári, I. Bársony: **Monitoring the tyre deformation on a vehicle on the run**, EUROSENSORS 2016, Budapest, Hungary



Piezoelectric ZnO nano-rods – for high-resolution fingerprint sensing

Previous work





- Bending and measurement in Atomic Force Microscope



Piezoelectric ZnO nano-rods – for high-resolution fingerprint sensing

Previous work





- Design and fabrication a 25x10 nano-array



Piezoelectric ZnO nano-rods – for high-resolution fingerprint sensing

Previous work




- Design and fabrication a 25x10 nano-array



Piezoelectric ZnO nano-rods – for high-resolution fingerprint sensing

Previous work




- Assembling a new measurement setup



Piezoelectric ZnO nano-rods – for high-resolution fingerprint sensing

Previous work


- Results




Piezoelectric ZnO nano-rods

Current work

- Measurement of the entire array in the same time

[illegible]

Piezoelectric ZnO nano-rods – for high-resolution fingerprint sensing



Piezoelectric ZnO nano-rods

Publications in this topic

Papers

- Bouvet-Marchand A, Graillot A, Volk J, Dauksevicius R, Sturm C, Grundmann M, Saoutieff E, Viana A, Christian B, Lebedev V, Rado J, Lukacs I E, N Q Khanh, Grosso D, Loubat C, **Design of UV-crosslinked polymeric thin layers for encapsulation of piezoelectric ZnO nanowires for pressure-based fingerprint sensors**, JOURNAL OF MATERIALS CHEMISTRY C in press: p. in press. (2018)
- Seifikar Masoud, Christian Björn P, Volk János, Radó János, Lukács István E, Dauksevicius Rolanas, Gaidys Rimvydas, Lebedev Vadim, Viana Antoine, O'Reilly Eoin P, **Direct observation of spontaneous polarization induced electron charge transfer in stressed ZnO nanorods**, NANO ENERGY 43: pp. 376-382. (2018)





Oral presentation:

- János Volk, István E. Lukács, Nguyen Quoc Khánh, János Radó, Róbert Erdélyi: **Bottom contacted piezoelectric nanowire arrays**, NGPT 2016, Rome, Italy
- J. Volk, J. Radó, I. E. Lukács, N. Q. Khánh, R. Erdélyi, G. Battistig, C. Sturm, M. Grundmann, A. Graillot, C. Loubat: **Integrated piezoelectric nanowire arrays for high resolution tactile mapping**, EUROSENSORS 2016, Budapest, Hungary

Piezoelectric AlN thin film for 3D force sensor




Previous work

- Test deposition with different parameters
- Qualification of complete thin films



Piezoelectric AlN thin film for 3D force sensor

Current work



Publications in this topic

Paper

Udvardi Péter, Radó János, Straszner András, Ferencz János, Hajnal Zoltán, Soleimani Saeedeh, Schneider Michael, Schmid Ulrich, Révész Péter, Volk János, **Spiral-Shaped Piezoelectric MEMS Cantilever Array for Fully Implantable Hearing Systems**, MICROMACHINES 8:(10) Paper 311. 13 p. (2017)

Further information

Completed courses

- Szilárdtest kémia
- Nanotechnológia
- Polimerek kémiája és fizikája
- Mikroelektronikai anyagok és szerkezetek vizsgálati módszerei
- Félvezető technológiák
- Ragasztás mentes szeletkötés
- Mikro és nano elektromechanikus szerkezetek

Participation in projects:

- Incite Eniac (Call 2013-1/621278-2)
- PiezoMat (grant no. 611019)
- KoFah (NVKP_16-1-2016-0018)

Further presentations

- Tapintásérzékelés az orvosi robotikában
(MTA Székház-Magyar Tudomány Ünnepe–Emberközpontú Technológia)
- MEMS technológiával előállított 3D erőmérő szenzorok
(Kandó Konferencia 2017)



***Thank you for your
attention!***